

	Company Name: NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	Document number: POL-GEN-QM-001
		Document Type: POL
TITLE: QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY		
Date of Issue: 2019-09-15	Issue No.: 10A	Page No.: Page 1 of 8

at

National Metrology Institute of Ethiopia
 B67, Street 1405, Woreda 06, Bole Sub city
 P.O. Box 5722
 Addis Ababa
 Ethiopia
 Phone:(+251 116) 671 551/ 452 046/460 542
 Fax:(+251 116) 459 312/463 028
 E-mail: info@nmie.net
 Website:www.nmie.net

Measured quantities:
 Conventional mass,
 temperature, Pressure,
 Volume, electrical quantities,
 and balance.

Head: Mr. Tadesse Gergiso, MSc Science of Measurement

Quality Manager Mr. Fikreab Markos, MSc Quality and Productivity Management

Deputy:

- Ms. Abeb Yitesha, BSc Chemistry
- Mr. Afework Gemed, BSc Physics
- Mr. Andargie Belete, BSc Chemistry
- Mr. Ayele Gashaw, BSc Chemistry
- Mr. Boja Beyi, Bsc Chemistry
- Mr. Birhan Tesfaye, BED Physics
- Mr. Dagmawi Woldetensai, Bsc Physics
- Mr. Elias Workneh, BSc Applied Physics
- Mr. Embaye Amine, BSc Chemistry
- Mr. Ermias Gizaw, BSc Applied Physics
- Mr. Fantahun Tamene, BSc Pysics
- Ms. Gelane Hailu, BSc Chemistry
- Mrs. Genet Nurga, MSc Chemistry
- Mr. Gizachew Betru, Bsc Chemistry
- Mrs. Hana Atsbaha, Bsc Chemistry
- Mr. Henok Kebede, Bsc Applied Physics
- Mr. Julius Dejene, BSc Physics
- Mr. Maymer Sergawi, Advanced General Mechanic technician
- Mr. Miliyon Yohans, MSc Material Science
- Mr. Misganaw Tsegaye, BSc Physics
- Mr. Mulugeta Derebew, Msc Measurement science
- Mr. Mulugeta Tesfaye, Bsc Physics
- Mr. Nesredin Nezir, Bsc Physics
- Mr. Nigussie Birhanu, Bsc Physics
- Mr. Ragasa Lamesa, BSc Physics
- Mr. Seid Fentaw, Bsc Applied Chemistry
- Mr. Solomon Assefa, Bsc Chemistry
- Mr. Wondwossen Birhanu, BSc Electrical Engineer

Accredited Since 2007-07-18

	Company Name: NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	Document number: POL-GEN-QM-001
		Document Type: POL
TITLE: QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY		
Date of Issue: 2019-09-15	Issue No.: 10A	Page No.: Page 2 of 8

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / Procedure ¹⁾	Best measurement capability ²⁾	Remarks
Conventional mass *)	1 mg, 2 mg, 5 mg	OIML R111-1-2004	0.006 mg	For weight pieces according to OIML recommendation R 111-1-2004, Class F ₁
	10 mg		0.008 mg	
	20 mg		0.010 mg	
	50 mg		0.012 mg	
	100 mg		0.016 mg	
	200 mg		0.020 mg	
	500 mg		0.025 mg	
	1 g		0.03 mg	
	2 g		0.04 mg	
	5 g		0.05 mg	
	10 g		0.06 mg	
	20 g		0.08 mg	
	50 g		0.10 mg	
	100 g		0.16 mg	
	200 g		0.3 mg	
	500 g		0.8 mg	
	1 kg		1.6 mg	
	2 kg		3.0 mg	
	5 kg		8.0 mg	
	10 kg		16 mg	
20 kg	30 mg			
Conventional mass *)	1 mg to 100 mg	OIML R111-1-2004	0.05 mg	For free nominal values
	> 100 mg to 200 mg		0.06 mg	
	> 200 mg to 500 mg		0.08 mg	
	> 500 mg to 1 g		0.10 mg	
	> 1 g to 2 g		0.12 mg	
	> 2 g to 5 g		0.16 mg	
	> 5 g to 10 g		0.20 mg	
	> 10 g to 20 g		0.25 mg	
	> 20 g to 50 g		0.30 mg	
	> 50 g to 100 g		0.5 mg	
	> 100 g to 10 kg		$5 \cdot 10^{-6} \cdot m_c$	
Temperature Resistance	-80 °C to 30 °C	Alcohol bath DAkKS-DKD-R 5-1:2010	20 mK	Alcohol bath, Water bath,

The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

	Company Name: NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	Document number: POL-GEN-QM-001
		Document Type: POL
TITLE: QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY		
Date of Issue: 2019-09-15	Issue No.: 10A	Page No.: Page 3 of 8

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / Procedure ¹⁾	Best measurement capability ²⁾	Remarks
thermometers *)	> 30 °C to 80 °C	Water bath DAkKS-DKD-R 5-1:2010	30 mK	Comparison with standard resistance thermometer
	> 80 °C to 200 °C	Oil bath DAkKS-DKD-R 5-1:2010	0.10 K	
	> 200 °C to 660 °C	Fluidized bath DAkKS-DKD-R 5-1:2010	0.30 K	
Thermocouples, noble metal *)	0 °C to 200 °C	Liquid bath DAkKS-DKD-R 5-3:2010	0,5 K	Comparison with resistance thermometer
	> 200 °C to 600 °C	Fluidized bath DAkKS-DKD-R 5-3:2010	0,7 K	
	> 600 °C to 900 °C	Furnace DAkKS-DKD-R 5-3:2010	1,1 K	Comparison with thermocouple
	> 900 °C to 1200 °C		1,5 K	
Thermocouples, base metal *)	0 °C to 200 °C	Furnace DAkKS-DKD-R 5-3:2010	1,0 K	Comparison with thermocouple
	> 200 °C to 600 °C		1,0 K	
	> 600 °C to 900 °C		1,5 K	
	> 900 °C to 1200 °C		2,0 K	
Liquid-in-glass thermometers	-80 °C to 30 °C	Alcohol bath PRO-PMT/TH-004 Revision 9A	25 mK	Comparison with standard resistance thermometer
	> 30 °C to 80 °C	Water bath PRO-PMT/TH-004 Revision 9A	30 mK	
	> 80 °C to 200 °C	Oil bath PRO-PMT/TH-004 Revision 9A	0.10 K	
	>200 °C to 360 °C	Fluidized bath PRO-PMT/TH-004 Revision 9A	0.30 K	
Direct reading electrical thermometers *)	-80 °C to 30 °C	Alcohol bath DAkKS-DKD-R 5-1:2010 DAkKS-DKD-R 5-3:2010	20 mK	Comparison with standard resistance thermometer
	> 30 °C to 80 °C	Water bath DAkKS-DKD-R 5-1:2010 DAkKS-DKD-R 5-3:2010	30 mK	
	> 80 °C to 200 °C	Oil bath DAkKS-DKD-R 5-1:2010 DAkKS-DKD-R 5-3:2010	0.10 K	
	> 200 °C to 660 °C	Fluidized bath DAkKS-DKD-R 5-1:2010 DAkKS-DKD-R 5-3:2010	0.30 K	
Mechanical (dial) thermometers	-40 °C to 80 °C	Alcohol or Water bath PRO-PMT/TH-005 Revision 9A	30 mK	Comparison with standard resistance thermometer;
	> 80 °C to 200 °C	Oil bath PRO-PMT/TH-005 Revision 9A	0.10 K	
	>200 °C to 400 °C	Fluidized bath PRO-PMT/TH-005 Revision 9A	0.30 K	

The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

	Company Name: NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	Document number: POL-GEN-QM-001
		Document Type: POL
TITLE: QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY		
Date of Issue: 2019-09-15	Issue No.: 10A	Page No.: Page 4 of 8

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / Procedure ¹⁾	Best measurement capability ²⁾	Remarks
Volume of liquids Piston operated pipettes *)	5 µL	ISO 8655:2002	0.026 µL	
	10 µL		0.04 µL	
	20 µL		0.08 µL	
	50 µL		0.14 µL	
	100 µL		0.3 µL	
	200 µL		0.6 µL	
	500 µL		1.2 µL	
	1000 µL		2.4 µL	
Volumetric burettes *)	1 mL	ISO 4787:2010	2.8 µL	
	2 mL		2.9 µL	
	5 mL		4.3 µL	
	10 mL		8.1 µL	
	25 mL		22 µL	
	50 mL		38 µL	
	100 mL		70 µL	
Volumetric pipettes *)	0.5 mL	ISO 4787:2010	1.5 µL	
	1 mL		4.4 µL	
	2 mL		5.6 µL	
	5 mL		7.4 µL	
	10 mL		8.7 µL	
	20 mL		12 µL	
	25 mL		16 µL	
	50 mL		17 µL	
	100 mL		39 µL	
	200 mL		46 µL	
Volumetric flask *)	1 mL, 2 mL, 5 mL, 10 mL	ISO 4787:2010	11 µL	
	20 mL, 25 mL		19 µL	
	50 mL		30 µL	
	100 mL		41 µL	
	200 mL		59 µL	
	250 mL		59 µL	
	500 mL		85 µL	
	1000 mL		0.13 mL	
	2000 mL		0.22 mL	
	5000 mL		0.39 mL	
Measuring cylinders *)	5 mL	ISO 4787:2010	35 µL	
	10 mL		61 µL	
	25 mL		93 µL	
	50 mL		0.13 mL	
	100 mL		0.20 mL	
	250 mL		0.36 mL	

The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

	Company Name:	NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	Document number:	POL-GEN-QM-001
			Document Type:	POL
TITLE:				
QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY				
Date of Issue: 2019-09-15		Issue No.: 10A		Page No.: Page 5 of 8

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / Procedure ¹⁾	Best measurement capability ²⁾	Remarks
	500 mL		0.58 mL	
	1000 mL		0.93 mL	
	2000 mL		1.5 mL	
Provers *)	5 L	ISO 4787:2010	2.1 mL	
	10 L		3.4 mL	
	20 L		3.8 mL	
DC voltage Zener standards	10 V	Direct comparison with Zener standard	30 μ V	
Sources, fixed values	0.1 V		1.2 μ V	
	1 V		3.0 μ V	
	10 V		45 μ V	
	100 V		0.50 mV	
	1000 V		12 mV	
Measuring instruments	10 mV to 11 V		$8.0 \cdot 10^{-6} \cdot U + 6.0 \mu\text{V}$	<i>U: measured value</i>
	> 11 V to 22 V		$9.0 \cdot 10^{-6} \cdot U + 4.0 \mu\text{V}$	
	> 22 V to 275 V		$10 \cdot 10^{-6} \cdot U + 90 \mu\text{V}$	
	> 275 V to 1 KV		$16 \cdot 10^{-6} \cdot U + 0.17 \text{ mV}$	
DC current Sources, fixed values	0.1 A		$0.65 \cdot 10^{-3} \cdot I$	<i>I: measured value</i>
	2 A		$0.60 \cdot 10^{-3} \cdot I$	
	4A, 6A, 8A, 10A, 12A, 14A, 16A, 18A, 20A		$0.50 \cdot 10^{-3} \cdot I$	
Measuring instruments	3.3 mA to 33 mA		$7.0 \cdot 10^{-6} \cdot I + 40 \mu\text{A}$	<i>I: measured value</i>
	>33 mA to 330 mA		$60 \cdot 10^{-6} \cdot I + 40 \mu\text{A}$	
	>330 mA to 1 A		$0.23 \cdot 10^{-3} \cdot I + 60 \mu\text{A}$	
	>1 A to 3 A		$0.45 \cdot 10^{-3} \cdot I + 50 \mu\text{A}$	
	>3 A to 11 A		$0.58 \cdot 10^{-3} \cdot I + 0.58 \mu\text{A}$	
	>11 A to 20 A		$1.1 \cdot 10^{-3} \cdot I + 0.90 \text{ mA}$	
DC resistance Resistors, fixed values	1 m Ω		$72 \cdot 10^{-6} \cdot R$	<i>R: measured value</i>
	10 m Ω		$28 \cdot 10^{-6} \cdot R$	
	100 m Ω		$22 \cdot 10^{-6} \cdot R$	
	1 Ω		$13 \cdot 10^{-6} \cdot R$	
	10 Ω		$13 \cdot 10^{-6} \cdot R$	
	100 Ω		$13 \cdot 10^{-6} \cdot R$	
	1 k Ω		$10 \cdot 10^{-6} \cdot R$	
	10 k Ω		$10 \cdot 10^{-6} \cdot R$	
	100 k Ω		$12 \cdot 10^{-6} \cdot R$	
	1 M Ω		$20 \cdot 10^{-6} \cdot R$	
DC resistance Measuring instruments	0.10 Ω to 11 Ω		$15 \cdot 10^{-6} \cdot R + 1.2 \text{ m}\Omega$	<i>R: measured value</i>
	> 11 Ω to 33 Ω		$41 \cdot 10^{-6} \cdot R + 20 \text{ m}\Omega$	
	> 33 Ω to 110 Ω		$13 \cdot 10^{-6} \cdot R + 20 \text{ m}\Omega$	

The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

	Company Name:	Document number:
	NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	
TITLE:		Document Type:
QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY		POL
Date of Issue: 2019-09-15	Issue No.: 10A	Page No.: Page 6 of 8

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / Procedure ¹⁾	Best measurement capability ²⁾	Remarks
	> 110 Ω to 330 Ω		$13 \cdot 10^{-6} \cdot R + 20 \text{ m } \Omega$	
	> 330 Ω to 1.1 kΩ		$13 \cdot 10^{-6} \cdot R + 20 \text{ m } \Omega$	
	>1.1 kΩ to 3.3 kΩ		$7.0 \cdot 10^{-6} \cdot R + 0.10 \text{ } \Omega$	
	>3.3 kΩ to 11 kΩ		$1.0 \cdot 10^{-6} \cdot R + 12 \text{ m } \Omega$	
	>11 kΩ to 33 kΩ		$15 \cdot 10^{-6} \cdot R + 0.80 \text{ } \Omega$	
	>33 kΩ to 110 kΩ		$7.0 \cdot 10^{-6} \cdot R + 0.60 \text{ } \Omega$	
	>110 kΩ to 330 kΩ		$10 \cdot 10^{-6} \cdot R + 1.5 \text{ } \Omega$	
	>330 kΩ to 1.1 MΩ		$7.0 \cdot 10^{-6} \cdot R + 12 \text{ } \Omega$	
	>1.1 MΩ to 3.3 MΩ		$70 \cdot 10^{-6} \cdot R + 0.17 \text{ k } \Omega$	
	>3.3 MΩ to 11 MΩ		$0.20 \cdot 10^{-3} \cdot R + 0.24 \text{ k } \Omega$	
	>11 MΩ to 33 MΩ		$0.37 \cdot 10^{-3} \cdot R + 2.6 \text{ k } \Omega$	
	>33 MΩ to 110 MΩ		$0.93 \cdot 10^{-3} \cdot R + 2.3 \text{ k } \Omega$	
	>110 MΩ to 330 MΩ		$3.5 \cdot 10^{-3} \cdot R + 0.12 \text{ M } \Omega$	
	>330 MΩ to 1.1 GΩ		$18 \cdot 10^{-3} \cdot R + 0.58 \text{ M } \Omega$	
AC voltage Measuring instruments	10 mV to 33 mV	40 Hz to 10 kHz	$4.0 \cdot 10^{-6} \cdot U + 0.50 \text{ mV}$	<i>U: measured value</i>
	>33 mV to 0.33 V		$16 \cdot 10^{-6} \cdot U + 0.50 \text{ mV}$	
	>0.33 V to 3.3 V	45 Hz to 10 kHz	$0.12 \cdot 10^{-3} \cdot U + 0.50 \text{ mV}$	
	>3.3 V to 33 V		$0.20 \cdot 10^{-3} \cdot U + 1.2 \text{ mV}$	
	>33 V to 330 V		$0.24 \cdot 10^{-3} \cdot U + 8.0 \text{ mV}$	
	>330 V to 1000 V		$0.33 \cdot 10^{-3} \cdot U + 70 \text{ mV}$	
AC current Measuring instruments	3.3 mA to 33 mA	45 Hz to 1 kHz	$20 \cdot 10^{-6} \cdot I + 0.24 \text{ mA}$	<i>I: measured value</i>
	>33 mA to 330 mA		$0.25 \cdot 10^{-3} \cdot I + 0.16 \text{ mA}$	
	>330 mA to 1 A		$0.60 \cdot 10^{-3} \cdot I + 0.12 \text{ mA}$	
	>1A to 3 A		$0.70 \cdot 10^{-3} \cdot I + 0.12 \text{ mA}$	
	>3A to 11 A		$0.70 \cdot 10^{-3} \cdot I + 3.0 \text{ mA}$	
	>11A to 20 A		$1.5 \cdot 10^{-3} \cdot I + 6.0 \text{ mA}$	
Pressure Absolute pressure p_{abs} *)	0.14 bar to 1.7 bar	DKD-R 6-1:2014 EURAMET Calibration Guide No. 17 Version 3.0 DIN EN 837:1997 EURAMET /cg-3/v.01	$2.7 \cdot 10^{-5} \cdot p_{abs} + 18 \text{ } \mu\text{bar}$	Pressure medium: Gas The uncertainty of the residual pressure has to be added
	>1.7 bar to 7.0 bar		$2.8 \cdot 10^{-5} \cdot p_{abs} + 22 \text{ } \mu\text{bar}$	
	>7.0 bar to 70 bar		$3.2 \cdot 10^{-5} \cdot p_{abs} + 0.10 \text{ mbar}$	
Gauge pressure p_e *)	0 bar	DKD-R 6-1:2014 EURAMET Calibration Guide No. 17 Version 3.0 DIN EN 837:1997 EURAMET /cg-3/v.01	$2.7 \cdot 10^{-5} \cdot p_e + 18 \text{ } \mu\text{bar}$	Pressure medium: Gas
	0.014 bar to 1.7 bar		$2.8 \cdot 10^{-5} \cdot p_e + 22 \text{ } \mu\text{bar}$	
	> 1.7 bar to 7.0 bar		$3.2 \cdot 10^{-5} \cdot p_e + 0,10 \text{ } \mu\text{bar}$	
	> 7.0 bar to 70 bar		25 mbar	
	> 70 bar to 140 bar			
	0 bar	DKD-R 6-1:2014 EURAMET Calibration Guide No. 17 Version 3.0 DIN EN 837:1997 EURAMET /cg-3/v.01	$7.0 \cdot 10^{-5} \cdot p_e + 0.15 \text{ mbar}$	Pressure medium: Oil
	0.5 bar to 60 bar		$1.4 \cdot 10^{-4} \cdot p_e + 5.8 \text{ mbar}$	
	> 60 bar to 1200 bar			

The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

	Company Name: NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	Document number: POL-GEN-QM-001
		Document Type: POL
TITLE: QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY		
Date of Issue: 2019-09-15	Issue No.: 10A	Page No.: Page 7 of 8

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / Procedure ¹⁾	Best measurement capability ²⁾	Remarks
Non-automatic electronic weighing instruments with digital indicator *)	Up to 100 g	EURAMET Calibration Guide No. 18, Version 4.0	$2 \cdot 10^{-6}$	with weights according to OIML R 111-1:2004, Class E ₂
	Up to 60 kg		$1 \cdot 10^{-5}$	with weights according to OIML R 111-1:2004, Class F ₁
	Up to 300 kg		$1 \cdot 10^{-4}$	with weights according to OIML R 111-1:2004, Class M ₁
Temperature Ovens and autoclaves (with air circulation) *)	25 °C to 400 °C	DAKKS-DKD-R 5-7:2010 EURAMET cg-20, Version 4.0	0.60 K	standard thermometers of 12 RTD, connected with one temperature measuring device and multiplexer
Climatic chambers (with air circulation) *)	-70 °C to 150 °C		0.50 K	
Calibration baths *)	-80 °C to 250 °C		0.15 K	
Calibration baths *)	-40 °C to 250 °C		0.50 K	

Mobile Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / Procedure ¹⁾	Best measurement capability ²⁾	Remarks
Conventional mass *)	1 g	OIML R 111-1:2004	0.1 mg	For weight pieces according to OIML recommendation R 111-1:2004, Class F ₂
	2 g			
	5 g			
	10 g			
	20 g			
	50 g			
	100 g			
	200 g			
	500 g			
	1 kg			
	2 kg			
	5 kg			
	10 kg			
	20 kg			
Gauge pressure pe *)	0 kPa to 400 kPa	DKD-R 6-1:2014 only sequence B and C	0.5 kPa	Pressure medium: Oil
	>400 kPa to 1MPa			
	>1MPa to 10 MPa			
	>10 MPa to 25 MPa			
	>25 MPa to 60 MPa			
			1.5	
			20 kPa	
			45 kPa	
			200 kPa	

The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

	Company Name: NATIONAL METROLOGY INSTITUTE OF ETHIOPIA LABORATORY MANAGEMENT SYSTEM	Document number: POL-GEN-QM-001
		Document Type: POL
TITLE: QUALITY MANUAL ANNEX 1: SCOPE OF ACCREDITATION OF NMIE CALIBRATION LABORATORY		
Date of Issue: 2019-09-15	Issue No.: 10A	Page No.: Page 8 of 8

Temperature Liquid-in-glass thermometers	-20 °C to 360 °C	Dry block calibrator PRO-MOB/TH-002, Version 1A	0.20 K	comparison with a standard resistance thermometer
Direct reading thermometers with resistance sensor *)	-20 °C to 660 °C	Dry block calibrator DAkKS-DKD-R 5-1:2010	0.20 K	
Direct reading thermometers with thermocouple sensor *)	-20 °C to 660 °C	Dry block calibrator DAkKS-DKD-R 5-3:2010	0.20 K	
Mechanical (dial) thermometers	-20 °C to 660 °C	Dry block calibrator PRO-MOB/TH-001, Version 1A	0.25 K	
Volume of liquids Volumetric burettes *)	10 mL	Gravimetric method ISO 4787:2010	7 µL	
	25 mL		10 µL	
	50 mL		22 µL	
	100 mL		37 µL	
Volumetric pipettes *)	10 mL	ISO 4787:2010	10 µL	
	20 mL		13 µL	
	25 mL		15 µL	
	50 mL		17 µL	
	100 mL		27 µL	
Volumetric flask *)	200 mL	ISO 4787:2010	34 µL	
	10 mL		22 µL	
	20 mL		37 µL	
	25 mL		37 µL	
	50 mL		55 µL	
	100 mL		78 µL	
	200 mL		0.14 mL	
	250 mL		0.14 mL	
	500 mL		0.21 mL	
	1000 mL		0.29 mL	
2000 mL	0.40 mL			
Measuring cylinders *)	5000 mL	ISO 4787:2010	0.79 mL	
	10 mL		89 µL	
	25 mL		0.17 mL	
	50 mL		0.27 mL	
	100 mL		0.41 mL	
	250 mL		0.73 mL	
	500 mL		1.2 mL	
	1000 mL		1.9 mL	
Provers *)	2000 mL	ISO 4787:2010	3.1 mL	
	5 L		1.5 mL	
	10 L		2.0 mL	
	20 L		4.1 mL	

The best measurement capabilities are stated according to EA-4/02. These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.